

(g) The package label or package insert shall state the blood group antigens that have been tested for and found present or absent on the cells of each donor, or refer to such information in an accompanying antigenic constitution matrix. Cells for ABO Serum Grouping are exempt from this requirement. The package insert or antigen constitution matrix shall list each of the antigens tested with only one source of antibody.

(h) The package label or package insert shall bear the cautionary statement: "The reactivity of the product may decrease during the dating period."

(i) The package insert of a product intended for the detection or identification of unexpected antibodies shall note that the rate at which antigen reactivity (e.g., agglutinability) is lost is partially dependent upon individual donor characteristics that are neither controlled nor predicted by the manufacturer.

(j) The package insert shall provide adequate directions for use.

(k) The package insert shall bear the statement:

"CAUTION: ALL BLOOD PRODUCTS SHOULD BE TREATED AS POTENTIALLY INFECTIOUS. SOURCE MATERIAL FROM WHICH THIS PRODUCT WAS DERIVED WAS FOUND NEGATIVE WHEN TESTED IN ACCORDANCE WITH CURRENT FDA REQUIRED TESTS. NO KNOWN TEST METHODS CAN OFFER ASSURANCE THAT PRODUCTS DERIVED FROM HUMAN BLOOD WILL NOT TRANSMIT INFECTIOUS AGENTS."

(l) The package insert or the antigenic constitution matrix for each lot of product shall specify the date of manufacture or the length of the dating period.

(m) Manufacturers shall identify with a permanent donor code in the product labeling each donor of peripheral blood used for detection or identification of unexpected antibodies.

[52 FR 37450, Oct. 7, 1987, as amended at 67 FR 9587, Mar. 4, 2002]

§ 660.36 Samples and protocols.

(a) The following shall be submitted to the Center for Biologics Evaluation and Research Sample Custodian (ATTN: HFM-672) (see mailing address-

es in § 600.2 of this chapter), within 30 days after each routine establishment inspection by FDA.

(1) From a lot of final product, samples from a cell panel intended for identification of unexpected antibodies. The sample shall be packaged as for distribution and shall have at least 14 days remaining in the dating period when shipped to the Center for Biologics Evaluation and Research.

(2) A protocol which shall include the following:

(i) Complete test records of at least two donors of the samples submitted, including original and confirmation phenotyping records.

(ii) Bleeding records or receipt records which indicate collection date, volume, and HBsAg test results.

(iii) Manufacturing records which document all steps involved in the preparation of the product.

(iv) Test results which verify that the final product meets specifications.

(v) Identity test results.

(b) A copy of the antigenic constitution matrix specifying the antigens present or absent shall be submitted to the Director, Center for Biologics Evaluation and Research, at the time of initial distribution of each lot of Reagent Red Blood Cells for detection or identification of unexpected antibodies. Products designed exclusively to identify Anti-A, Anti-A₁, and Anti-B, as well as products composed entirely of umbilical cord cells, are excluded from this requirement.

(c) Except for umbilical cord samples, whenever a new donor is used, a sample of red blood cells from each new donor used in a cell panel intended for the identification of unexpected antibodies shall be submitted by the manufacturer to the Director, Center for Biologics Evaluation and Research. The sample should contain a minimum volume of 0.5 milliliter of red blood cells.

[52 FR 37450, Oct. 7, 1987, as amended at 55 FR 11013 and 11015, Mar. 26, 1990; 67 FR 9587, Mar. 4, 2002; 70 FR 14985, Mar. 24, 2005]

Subpart E—Hepatitis B Surface Antigen

SOURCE: 44 FR 36382, June 22, 1979, unless otherwise noted.